



Certificate of Compliance

Certificate: 1746881 (LR 64969)

Master Contract: 167534

Project: 2722508

Date Issued: May 6, 2014

Issued to: Mine Safety Appliances Company
North American Division
1000 Cranberry Woods Dr
Cranberry Township, PA 16066-5296
USA
Attention: Semyon (Sam) Baum

The products listed below are eligible to bear the CSA Mark shown



Glenn Black

Issued by: Glenn Black

PRODUCTS

CLASS 4828 01 - SIGNAL APPLIANCES - - Combustible Gas Detection Instruments-For Hazardous Locations

CLASS 4828 02 - SIGNAL APPLIANCES - Toxic Gas Detection Instruments - For Hazardous Locations

Class 4828 01

Class I, Groups A, B, C and D:

“Ultima XE Gas Monitor” consisting of an “Ultima XE Main” with integral “Ultima XE Sensor”, input rated 7-30V dc, 1A max., 4-20mA signal output, with optional internal relays rated 5A @ 250V ac or 30V dc. Must be used with Weatherguard P/N 10028904. May be used with CSA certified “Ultima” Calibrator and CSA certified “Ultima” Controller.

“Ultima XE Gas Monitor” System consisting of an “Ultima XE Main” input rated 7-30V dc, 1A max., 4-20mA signal output, with optional internal relays rated 5A @ 250V ac or 30V dc, for connection to a “Ultima XE Remote Sensor” with integral “Ultima XE Sensor”. Must be used with Weatherguard P/N 10028904. May be used with CSA certified “Ultima” Calibrator and CSA certified “Ultima” Controller.

“Ultima X Gas Monitor” System consisting of an “Ultima XE Main with X3 Technology”, input rated 7-30V dc, 10 W max., with optional internal relays rated 5A @ 250V ac or 30V dc; for use with up to 3 of the following: integral Ultima XE Sensor, integral Ultima XIR Gas Monitor (Class I, Groups B, C and D only), Ultima XE Remote Sensor, Ultima XIR Remote Sensor (Class I, Groups B, C and D only), Ultima XE Remote Sensor with



Certificate: 1746881 (LR 64969)

Master Contract: 167534

Project: 2722508

Date Issued: May 6, 2014

AC Power Supply, input rated 120/240 V ac, 5 W max, Ultima XE Remote Sensor with DC Power Supply, input rated 9-30 V dc, 5 W max, Ultima XIR Remote Sensor with AC Power Supply(Class I, Groups B, C and D only), input rated 120/240 V ac, 5 W max, Ultima XIR Remote Sensor with DC Power Supply(Class I, Groups B, C and D only), input rated 9-30 V dc, 5 W max. Ultima XE Sensor” and “Ultima XE Remote Sensor” must be used with Weatherguard P/N 10028904. Environmental Guard p/n 10041265 may be used with Ultima XIR Gas Monitor and XIR remote sensor. May be used with CSA certified “Ultima” Calibrator and CSA certified “Ultima” Controller.

Class I, Groups B, C and D:

“Ultima XE Gas Monitor” consisting of an “Ultima XE Main” with integral “Ultima XIR Gas Monitor”, input rated 7-30V dc, 1A max., 4-20mA signal output, with optional internal relays rated 5A @ 250V ac or 30V dc. May be used with CSA certified “Ultima” Calibrator, CSA certified “Ultima” Controller and Environmental Guard p/n 10041265.

“Ultima XE Gas Monitor” System consisting of an “Ultima XE Main” input rated 7-30V dc, 1A max., 4-20mA signal output, with optional internal relays rated 5A @ 250V ac or 30V dc, for connection to a “Ultima XIR Remote Sensor” with integral “Ultima XIR Gas Monitor”. May be used with CSA certified “Ultima” Calibrator, CSA certified “Ultima” Controller and Environmental Guard p/n 10041265.

“Ultima XI Gas Monitor”, input rated 7-30V dc, 800mA max., 4-20mA signal output. May be used with a MSA environmental Guard p/n 10041265.

CLASS 4828 02

Class I, Groups A, B, C and D:

“Ultima XE Gas Monitor” consisting of a “Ultima XE Main” with integral “Ultima XE Sensor”, input rated 7-30V dc, 1A max., 4-20mA signal output, with optional internal relays rated 5A @ 250V ac or 30V dc. Must be used with Weatherguard P/N 10028904. May be used with CSA certified “Ultima” Calibrator and CSA certified “Ultima” Controller.

“Ultima XE Gas Monitor” System consisting of a “Ultima XE Main input rated 7-30V dc, 1A max., 4-20mA signal output, with optional internal relays rated 5A @ 250V ac or 30V dc, for connection to a “Ultima XE Remote Sensor” with integral “Ultima XE Sensor”. Must be used with Weatherguard P/N 10028904. May be used with CSA certified “Ultima” Calibrator and CSA certified “Ultima” Controller.

“Ultima X Gas Monitor” System consisting of an “Ultima XE Main with X3 Technology”, input rated 7-30V dc, 10 W max., with optional internal relays rated 5A @ 250V ac or 30V dc; for use with up to 3 of the following: integral Ultima XE Sensor, Ultima XE Remote Sensor, Ultima XE Remote Sensor with AC Power Supply, input rated 120/240 V ac, 5 W max, Ultima XE Remote Sensor with DC Power Supply, input rated 9-30 V dc, 5 W max. Ultima XE Sensor” and “Ultima XE Remote Sensor” must be used with Weatherguard P/N 10028904. May be used with CSA certified “Ultima” Calibrator and CSA certified “Ultima” Controller.



Certificate: 1746881 (LR 64969)

Master Contract: 167534

Project: 2722508

Date Issued: May 6, 2014

APPLICABLE REQUIREMENTS

CAN/CSA C22.2 No. 0-10 - General Requirements – Canadian Electrical Code, Part II

CSA Std C22.2 No. 30-M1986 - Explosion-Proof Enclosures for Use in Class I Hazardous Locations

CSA Std C22.2 No. 142-M1987 - Process Control Equipment

CSA Std C22.2 No. 152-M1984 - Combustible Gas Detection Instruments

CSA Std C22.2 No. 157-92 - Intrinsically Safe and Non-Incendive Equipment for Use in Hazardous Locations