

Addendum A

# Standards Compliance Certifications For Altair® 4XR Operating Manual



Order No.: 10175895/05  
CR 800000023860

### Device Cleaning

Cleaning of the apparatus is recommended following environmental exposure, to ensure optimal performance:

- **Routine Cleaning:** Clean the exterior of the device regularly using only a damp cloth. Do not use cleaning agents, as many contain silicone, which will damage the combustible sensor.
- **Dust and Dirt Exposure:** Use a dry, soft bristled brush to remove any dust or dirt that has accumulated on the apparatus, especially at the sensor openings. If there is a buildup of dust or dirt particles remaining in the sensor area after brushing, use a vacuum to remove remaining particles, but maintain at least a 1/2 inch (1,2 cm) gap between the vacuum inlet and the apparatus.
- **Chemical Exposure:** If the equipment is likely to come into contact with aggressive substances, e.g. acidic liquids or gases that may attack metals or solvents that may affect polymeric materials, then it is the responsibility of the user to take suitable precautions that prevent it from being adversely affected thus ensuring that the type of protection is not compromised.
- **Water Exposure:** If the device is exposed to water, turn the device sensor side down and gently shake water off the sensor area. Any remaining water can be removed with a clean dry cloth. In the event that the device is immersed in water, allow the sensor inlets time to dry before retesting and returning to service. Drying time is dependent upon humidity conditions and the duration of immersion.

### Bluetooth SIG Statement

The design is listed as "Industrial Portable Gas Monitoring Equipment", Declaration ID D026835

[https://www.bluetooth.org/tpg/QLI\\_viewQDL.cfm?qid=26835](https://www.bluetooth.org/tpg/QLI_viewQDL.cfm?qid=26835)

#### Panasonic Bluetooth radio module certified to:

- FCC Part 15, FCC Identifier - T7V1316
- Industry Canada compliance to RSS-210. Industry Canada license - 216Q-1316

### Special Conditions for Safe Usage

#### WARNING!

#### Special Conditions for Safe Usage

- ▶ In the event of combustible sensor overrange, the device will enter a Lock Alarm state, which must be reset in a fresh air environment. To reset this alarm, cycle power off and on, in fresh air. Keep the device in the fresh air environment until LEL or CH4 readings have stabilized and then follow the Fresh Air Setup and Zero Calibration instructions contained in the user instruction manual
  - ▶ The RF radiation power used to activate the RFID tag antenna shall not exceed 6 Watts for EPL Group I applications or 2 Watts for EPL Group IIC applications.
-

## Certifications and Markings

See device label on your specific device, for applicable certification markings. The following label examples are for demonstration purposes only and may not accurately depict the current product certification status.

Information common to all labels:

Manufacturer:	MSA THE SAFETY COMPANY 1000 Cranberry Woods Drive Cranberry Township, PA 16066 USA
Product:	ALTAIR 4XR Gas Detector
Serial Number and Date:	XXXXXX = Serial Number DDDDD = Date Code

### North American (USA and Canada) Certifications and Markings

Agency: CSA Group

Permitted Hazardous Locations

#### USA

Class I, Division 1, Groups A, B, C, & D, Class II, Division 1 Groups E, F, & G Class III; T4

Class I, Zone 0, AEx ia IIC T4 Ga (Zone 0) - Without XCell Ex Sensor

Class I, Zone 0, AEx da ia IIC T4 Ga (Zone 0) - With XCell Ex Sensor

#### CANADA

Class I, Division 1, Groups A, B, C, & D, Class II, Division 1 Groups E, F, & G, Class III; T4

Ex ia IIC T4 Ga (Zone 0) - Without XCell Ex Sensor

Ex da ia IIC T4 Ga (Zone 0) - With XCell Ex Sensor

#### USA and CANADA

Ambient Temperature Range:  $-40\text{ }^{\circ}\text{C} \leq T_a \leq 60\text{ }^{\circ}\text{C}$ , Temperature code: T4

Ambient Temperature Range (Combustible Gas Performance):  $-20\text{ }^{\circ}\text{C} \leq T_a \leq 54\text{ }^{\circ}\text{C}$

Pollution Degree: 2, Altitude: 2000 m

## Certifications and Markings

Markings contained on label:



### WARNING : AVERTISSEMENT

SUBSTITUTION OF COMPONENTS MAY IMPAIR INTRINSIC SAFETY

USE ONLY BATTERY PACK P/N 10083913

BATTERIES MUST ONLY BE CHANGED OR CHARGED IN AN AREA KNOWN TO BE NONHAZARDOUS

UNDERSTAND MANUAL BEFORE OPERATING

ONLY THE COMBUSTIBLE GAS DETECTION PORTION OF THIS INSTRUMENT HAS BEEN ASSESSED FOR PERFORMANCE

NOT PERFORMANCE APPROVED FOR CL. II, CL III. SENSOR MAY BECOME CLOGGED AND NOT DETECT GAS OR WARN USER OF INABILITY

AVERTISSEMENT: LA SUBSTITUTION DE COMPOSANTS PEUT COMPROMETTRE LA SÉCURITÉ INTRINSÈQUE

UTILISEZ UNIQUEMENT BATTERIE P/N 10083913

NE CHANGER OU CHARGER LES BATTERIES QUE DANS DES EMPLACEMENTS DESIGNES NON DANGEREUX

LIRE ATTENTIVEMENT LES INSTRUCTIONS AVANT DE METTRE EN MARCHÉ

SEULEMENT LA PARTIE DE DÉTECTION DE GAZ COMBUSTIBLE DE CET INSTRUMENT A ÉTÉ ÉVALUÉE POUR PERFORMANCE

Intrinsic Safety Standards:

- CSA C22.2 No. 157,
- UL 913, and

Performance Standards:

- CSA C22.2 No. 152, and
- ANSI/ISA 60079-29-1
- Gas measurement range: 0 - 100% LEL Methane

**ATEX Certifications and Markings**

**EC-Type Examination Certificate:  
Sira 16ATEX2292**

Intrinsic Safety Standards:

- EN 60079-0:2012/A11:2013,
- EN 60079-1:2014, and
- EN 60079-11:2012

Ambient Temperature Range:

-40 °C ≤ T<sub>a</sub> ≤ 60 °C, Temperature code: T4

Combustible Performance Standards:

- EN 60079-29-1:2016

Ambient Temperature Range (Combustible Gas Performance): -20 °C ≤ T<sub>a</sub> ≤ 54 °C

**NOTE:** EN 60079-29-1 Combustible Gas Performance testing is for Group II only

Toxic Performance Standards:

- EN 45544-1:2015, EN 45544-3 (H<sub>2</sub>S)

Ambient Temperature Range (Toxic Gas Performance): -20 °C ≤ T<sub>a</sub> ≤ 60 °C

**NOTE:** TWA function for H<sub>2</sub>S is not supported

- EN 45544-1:2015, EN 45544-2:2015 (CO)

Ambient Temperature Range (Toxic Gas Performance): -20 °C ≤ T<sub>a</sub> ≤ 50 °C

Oxygen Performance Standards:

- EN 50104:2010 (O<sub>2</sub>)

Ambient Temperature Range (Oxygen Gas Performance): -20 °C ≤ T<sub>a</sub> ≤ 60 °C



### IEC Certifications and Markings

**Certificate Number: IECEx SIR 16.0096**

Intrinsic Safety Standards:

- IEC 60079-0 Edition 6.0,
- IEC 60079-1 Edition 7.0, and
- IEC 60079-11 Edition 6.0

Ambient Temperature Range:

$-40\text{ °C} \leq T_a \leq 60\text{ °C}$ , Temperature code: T4

Combustible Performance Standards:

- IEC 60079-29-1 Edition 1.0

**NOTE:** EN 60079-29-1 Combustible Gas Performance testing is for Group II only

Ambient Temperature Range (Combustible Gas Performance):  $-20\text{ °C} \leq T_a \leq 54\text{ °C}$



### Brazil (InMetro) Certifications and Markings

Certificate Number: NCC 17.0186 Intrinsic Safety Standards:

- ABNT NBR IEC 60079-0,
- ABNT NBR IEC 60079-1
- ABNT NBR IEC 60079-11

Marking:

Ex ia IIC T4 Ga (sem sensor XCell)

Ex da ia IIC T4 Ga (com sensor XCell)

$(-40\text{ °C} \leq T_{amb} \leq +60\text{ °C})$



06077-16-02640

“Este equipamento opera em caráter secundário, isto é, não tem direito a proteção contra interferência prejudicial, mesmo de estações do mesmo tipo, e não pode causar interferência a sistemas operando em caráter primário.”

Para maiores informações, consulte o site da ANATEL ([www.anatel.gov.br](http://www.anatel.gov.br))

For local MSA contacts, please visit us at **[MSAsafety.com](https://www.MSA.com)**

*Because every life has a **purpose...***