Australian/New Zealand

Certification Scheme for

EXPLOSION-PROTECTED ELECTRICAL EQUIPMENT

ANZEx Scheme

Certificate of Conformity

Certificate No.: ANZEx 05.	3045X	Issue No.: 5	Date of Issue: 03/08/2012
		Issue No.: 0	Date of Issue: 20/12/2005
Applicant:	1000 Cran Cranberry	ty Appliances Co. (Instrun berry Woods Drive Township PA 16066 – 520 ttes of America	
Electrical Apparatus:	Altair Port	table Single Gas Detector	
Type of Protection:	Ex ia		
Marking Code:	Ex ia IIC 7	$\Gamma 4 (Ta = -40 \ ^{\circ}C \ to \ 50 \ ^{\circ}C)$	
Manufacturer:	1000 Cran Cranberry	ty Appliances Co. (Instrun berry Woods Drive Township PA 16066 – 520 ttes of America	
Manufacturing Location(s):	As above		

The EPEE certification database located at http://www.anzex.com.au shows the validity of this Certificate.



This certificate and schedule shall not be reproduced except in full

Australian/New Zealand Certification Scheme for EXPLOSION-PROTECTED ELECTRICAL EQUIPMENT ANZEx Scheme Certificate of Conformity

Certificate No.: ANZEx 05.3045X	Issue No.: 5	Date of Issue: 03/08/2012
	Issue No.: 0	Date of Issue: 20/12/2005

This certificate is granted subject to the conditions as set out in Standards Australia/Standards New Zealand Miscellaneous Publication MP87.1:2008.

STANDARDS:

The electrical apparatus and any acceptable variations to it specified in the schedule of this certificate and the identified documents, was found to comply with the following standards:

IEC 60079-0: 2004 Electrical apparatus for explosive gas atmospheres Part 0: General requirements

IEC 60079-11:1999 Electrical apparatus for explosive gas atmospheres Part 11: Intrinsic safety 'i'

This Certificate **does not** indicate compliance with electrical safety and performance requirements other than those expressly included in the Standard(s) listed above.

ASSESSMENT & TEST REPORTS:

The equipment listed has successfully met the assessment and test requirements as recorded in:

Test Report No. and Issuing Body: Quality Assessment Report No. and Issuing Body: 33797, TestSafe Australia FR/INE/QAR08.0011/03, Ineris

File Reference:

2012/012105

Ujen Singh

03/08/2012

Date of Issue

Signed for and on behalf of issuing body

Quality & Certification Manager

Position

This certificate is not transferable and remains the property of the issuing body and must be returned in the event of it being revoked or not renewed.

This certificate and schedule shall not be reproduced except in full

Australian/New Zealand Certification Scheme for EXPLOSION-PROTECTED ELECTRICAL EQUIPMENT ANZEx Scheme

Certificate of Conformity

Certificate No.: ANZEx 05.3045X	Issue No.: 5	Date of Issue: 03/08/2012
	Issue No.: 0	Date of Issue: 20/12/2005

Schedule

EQUIPMENT:

The "Altair" portable single gas detector is initially available in three gas models e.g. O_2 , CO and H_2S . The O_2 model is called "Altair O_2 ". The gas sensing cells are of the electro-chemical type. The instrument is powered by a non-replaceable lithium cell. The LCD display panel on the front displays the months of battery life remaining or the alarm range when in alarm mode. A push button allows users to select the information viewed on the LCD display. Audible and visual alarm indicators are located on the front of the apparatus, and an imbalanced motor produces an optional vibration alert when in alarm mode.

CONDITIONS OF CERTIFICATION:

It is a condition of safe use that the lithium batteries of XENO XL-055F, Saft LST 17330 and EEMB ER 17335 shall be used in the apparatus.

Document	Sheets	Document Title	Issue	Date
No.				(yyyy-mm-dd)
SK3098-1013	1 of 7	SK3098_Altair Approval Critical Items (Labels)	Orig.	2005-12-15
SK3098-1013	2 of 7	SK3098_Altair Approval Critical Items (Assembly)	Orig.	2005-12-15
SK3098-1013	3 of 7	SK3098_Altair Approval Critical Items (Schematic)	Orig.	2005-12-15
SK3098-1013	4 of 7	SK3098_Altair Approval Critical Items (Parts List)	Orig.	2005-12-15
SK3098-1013	5 of 7	SK3098_Altair Approval Critical Items (Artworks)	Orig.	2005-12-15
SK3098-1013	6 of 7	SK3098_Altair Approval Critical Items (<i>Blank</i> page)	Orig.	2005-12-15
SK3098-1013	7 of 7	SK3098_Altair Approval Critical Items (Approval List)	Orig.	2005-12-15

DOCUMENTS:

Australian/New Zealand Certification Scheme for EXPLOSION-PROTECTED ELECTRICAL EQUIPMENT

ANZEx Scheme

Certificate of Conformity

Certificate No.: ANZEx 05.3045X	Issue No.: 5	Date of Issue: 03/08/2012
	Issue No.: 0	Date of Issue: 20/12/2005

Schedule of Variations

Variations Permitted by Issue 1:

 \triangleright Changes H₂S sensor part numbers.

These changes to the sensor part numbers do not affect the previous assessment in Test Report 27040. No further assessment is required.

Condition of Certification Relating to Issue 1:

The conditions remain as specified for Issue 0 of this certificate.

Drawings Relating to Issue 1:

Document No.	Sheets	Document Title	Issue	Date (yyyy-mm-dd)
SK3098-1013	7	Approval Critical Items, TestSafe I.S. (ANZEx/IECEx), Altair	1	2007-05-08
Altair.Parts.Details. 051222a119.wFoot er.doc	3 of 5	Parts Details, Altair	-	-

Variations Permitted by Issue 2:

Adding an alternate case overmould material RTP 2099E X 118617 B (Black) to the apparatus.

This change has been assessed in Test Report 30309.

Condition of Certification Relating to Issue 2:

The conditions remain as specified for Issue 0 of this certificate.

Drawings Relating to Issue 1:

Document No.	Sheets	Document Title	Issue	Date
SK3098-1013	6 of 8	Special Component Requirements	2	(yyyy-mm-dd) 2008-05-06

Australian/New Zealand Certification Scheme for EXPLOSION-PROTECTED ELECTRICAL EQUIPMENT

ANZEx Scheme

Certificate of Conformity

Certificate No.: ANZEx 05.3045X	Issue No.: 5	Date of Issue: 03/08/2012
	Issue No.: 0	Date of Issue: 20/12/2005

Variations Permitted by Issue 3:

- > The minimum ambient temperature changed from -20 °C to -40 °C.
- Removed sensor current parameter.
- Some text added/replaced for clarification.
- Added alternate static dissipative case overmould material RTP2099 EX 118617 B (various colours) to the apparatus.

These changes have been assessed in Test Report 31213.

Condition of Certification Relating to Issue 3:

The conditions remain as specified for Issue 0 of this certificate.

Drawings Relating to Issue 3:

Document No.	Sheets	Document Title	Issue	Date
				(yyyy-mm-dd)
SK3098-1013	8	Altair_TestSafe (ANZEx / IECEx)	3	2008-12-02

Variations Permitted by Issue 4:

> Only the revision number was changed to indicate a solder mask change.

These changes have been assessed in Test Report 33116.

> The QAR Issuer was changed from CSA to Ineris.

Condition of Certification Relating to Issue 4:

The conditions remain as specified for Issue 0 of this certificate.

Drawings Relating to Issue 4:

Document No.	Sheets	Document Title	Issue	Date
				(yyyy-mm-dd)
SK3098-1013	8	Altair_TestSafe (ANZEx / IECEx)	4	2011-03-09

Australian/New Zealand Certification Scheme for EXPLOSION-PROTECTED ELECTRICAL EQUIPMENT ANZEx Scheme

Certificate of Conformity

Certificate No.: ANZEx 05.3045X	Issue No.: 5	Date of Issue: 03/08/2012
	Issue No.: 0	Date of Issue: 20/12/2005

Variations Permitted by Issue 5:

- 1. The IR Transceiver (component designator U2) Sharp GP2W0110YP0F is being replaced by a similar device Vishay TFBS4650.
- 2. Associated capacitor C8 decreased in nominal capacitance from $1.0 \ \mu\text{F}$ to $0.1 \ \mu\text{F}$.
- 3. Protective resistor R3 previously 39.2 Ω nominal, ±1% tolerance, 0.5 W rating, has changed to 100 Ω nominal, ±1% tolerance, 0.5 W rating.

These changes have been assessed in Test Report 33797.

Condition of Certification Relating to Issue 5:

The conditions remain as specified for Issue 0 of this certificate.

Drawings Relating to Issue 5:

Document No.	Sheets	Document Title	Issue	Date
				(yyyy-mm-dd)
SK3098-1013	26	Altair / Altair O2 TestSafe	5	2012-05-18