

Physical Technical Testing Institute, Ostrava-Radvanice National testing authority No 210

Certification body No. 3051 accredited by CAI Praha

Supplement No. 8 to Certificate No. FTZÚ 09 E 0026

about verification of workplace atmospheres - Electrical apparatus used for the direct detection and direct concentration measurement of toxic gases and vapours

This certificate is issued for:

ALTAIR 5 Multigas Detector

Manufacturer: MSA - The Safety Company

Address:

1000 Cranberry Woods Drive, Cranberry Twp, PA 16066, USA

This supplement to certificate is valid for:

new performance tests for model ALTAIR 5X

change of manufacturer's name

Modification of certified apparatus and any of its approved variants are specified in documentation, list of which is mentioned in schedule of this supplement to the certificate.

The examination and test results are recorded in confidential Report No.

09E0026/8 dated 21.04.2017

FTZÚ - National testing authority No. 210 confirms, that product comply with requirements following standards:

EN 45544-1:2015, EN 45544-3:2015, EN 50104:2010, EN 50271:2010

Manufacturer (or applicant) listed in this supplement to the certificate is responsible for product conformity assurance in accordance with its specification (documentation) listed in schedule of this supplement to the certificate and for successful performance of all specified routine tests and verification.

This certificate is valid till:

04.04.2019

Responsible person:

Dipl. Ing. Lukáš Martinák Head of Certification Body

Date of issue, 21 04.2017

Page: 1/2

This supplement to certificate is granted subject to the general conditions of the FTZÚ, s.p. This supplement to certificate may only be reproduced in its entirety and without any change, schedule included.



Physical Technical Testing Institute, Ostrava-Radvanice National testing authority No 210

Certification body No. 3051 accredited by CAI Praha

Schedule

Supplement No. 8 to Certificate No. FTZÚ 09 E 0026

Product description changes:

Variation eight, to permit:

- a) The performance tests have been carried out according to standards EN 45544-1:2015 and EN 45544-3:2015.
- b) The name of the manufacturer has been changed. The previous name Mine Safety Appliances Company has been superseded by MSA-The Safety Company.
- c) The firmware has been updated to version v2.17.1.

Parameters recapitulation:

Performance tests have been carried out for:

- Gas H₂S in range 0-200ppm by electrochemical sensor. The TWA function is not supported.
- Gas O₂ in range 0-25% (v/v) by electrochemical sensor. The warm-up time is up to 180 sec.

List of documentation:

| Document No.: | Revision: | Date: | Pages: | Title: |
|---------------|-----------|------------|----------|---------------------------------------|
| 10116951 | 1_ | 2017 | 57 to 62 | Operating Manual – chapter 6 |
| SK3098-1184 | 8 | 16.03.2015 | 97 | FTZU approvals, ALTAIR5X / ALTAIR5XIR |
| SK3098-1307 | 03 | 10.07.2015 | 17 | Safety Discussion |

Responsible person:

Dipl. Ing. Lukáš Martinák Head of Certification Body Date of issue: 21.04.2017

Page: 2/2

This supplement to certificate is granted subject to the general conditions of the FTZÚ, s.p. This supplement to certificate may only be reproduced in its entirety and without any change, schedule included.