

FM Approvals LLC
1151 Boston-Providence Turnpike
Norwood, MA 02062
United States of America



ANNEX to IECEx FMG 07.0006X Issue No. 4

Marking

Ultima XL with XP Port:

Ex d [ib] IIC T6 Ta = -40 °C to +60 °C IP66
Ex nA [ib] IIC T6 Ta = -40 °C to +60 °C IP66
IECEX FMG 07.0006X

Ultima XL without XP Port:

Ex d IIC T6 Ta = -40 °C to +60 °C IP66
Ex nA IIC T6 Ta = -40 °C to +60 °C IP66
IECEX FMG 07.0006X

Ultima XE Main with XP Port:

Ex d [ib] IIC T4 Ta = -40 °C to +60 °C IP66
IECEX FMG 07.0006X

Ultima XE Main without XP Port:

Ex d IIC T4 Ta = -40 °C to +60 °C IP66
IECEX FMG 07.0006X

Ultima XE Sensor:

Ex d IIC T4 Ta = -40 °C to +60 °C IP66
IECEX FMG 07.0006X

Ultima XI Sensor:

Ex d IIC T5 Ta = -40 °C to +60 °C IP66
IECEX FMG 07.0006X

Ultima XIR Sensor:

Ex d IIC T5 Ta = -40 °C to +60 °C IP66
IECEX FMG 07.0006X

HART Module:

Ex d IIC T5 Ta = -40 °C to +60 °C IP66
IECEX FMG 07.0006X

Ultima X Power Supply:

Ex d IIC T4 Ta = -40 °C to +60 °C IP66
IECEX FMG 07.0006X

Ultima Calibrator

Ex ia IIC T3
IECEX FMG 07.0006X

Ultima Controller:

Ex ia IIC T3
IECEX FMG 07.0006X