

Operating Manual

# Multitest plus Console HP 230 V



MSA AUER GmbH  
Thiemannstrasse 1  
D-12059 Berlin

Germany

© MSA AUER GmbH. All rights reserved



## Declaration of Conformity

The manufacturer or his in the community established authorized representative

MSA AUER GmbH  
Thiemannstrasse 1  
D-12059 Berlin

declares that the product:

**Multitest plus Console HP 230 V**

is in conformance with the EMC Directive 2004/108/EC.

EN 61000-6-2:2005 and EN 61000-6-3:2007

We further declare that the product complies with the provisions of LVD Directive 2006/95/EC, with the following harmonized norms or normative documentation:

EN 61010-1:2002-08

*Dr. A. Schubert*

MSA AUER GmbH  
Dr. Axel Schubert  
R&D Instruments

Berlin, May 2009

**Contents**

<b>1 Safety Regulations .....</b>	<b>6</b>
1.1 Correct Use .....	6
1.2 Liability Information .....	6
1.3 Safety and Precautionary Measures .....	7
<b>2 Description .....</b>	<b>8</b>
2.1 Scope of Delivery .....	8
2.2 Overview .....	8
2.3 Operating Elements .....	9
2.4 Test Head .....	10
2.5 Test Head Maintenance .....	11
<b>3 Use .....</b>	<b>11</b>
3.1 Device Test .....	11
3.2 Activation .....	11
<b>4 Testing Negative Pressure Devices .....</b>	<b>12</b>
4.1 Lung Governed Demand Valve – Tightness Test without Medium Pressure .....	12
4.2 Lung Governed Demand Valve – Tightness Test with Medium Pressure .....	13
4.3 Lung Governed Demand Valve – Opening Pressure .....	13
4.4 Medium Pressure – Tightness Test .....	14
4.5 High Pressure – Tightness Test .....	14
4.6 High Pressure - Pressure and Gauge Accuracy Test .....	15
4.7 Mask - Tightness Test .....	15
<b>5 Testing Positive Pressure Devices .....</b>	<b>16</b>
5.1 Lung Governed Demand Valve – Tightness Test without Medium Pressure .....	16
5.2 Lung Governed Demand Valve – Tightness Test with Medium Pressure .....	17
5.3 Lung Governed Demand Valve – Closing Pressure .....	18
5.4 Medium Pressure – Tightness Test .....	19
5.5 High Pressure – Tightness Test .....	20
5.6 High Pressure - Pressure and Gauge Accuracy Test .....	20
5.7 Mask - Tightness Test .....	21
5.8 Mask - Opening Pressure of Exhalation Valve .....	21
<b>6 Testing Chemical Protective Suit .....</b>	<b>22</b>
6.1 Preparation .....	22
6.2 Stabilising the Chemical Protective Suit .....	22
6.3 Chemical Protective Suit – Tightness Test .....	23
6.4 Chemical Protective Suit – Valve Test .....	23

<b>7</b>	<b>Test Criteria for MSA Respiratory Protection Apparatus .....</b>	<b>24</b>
<b>8</b>	<b>Technical Data .....</b>	<b>25</b>
<b>9</b>	<b>Ordering Information .....</b>	<b>26</b>
9.1	Transition Adapters Mask / Lung Governed Demand Valves .....	26
9.2	Transition Adapter for Chemical Protective Suit .....	27
9.3	Accessories .....	28

## 1 Safety Regulations

### 1.1 Correct Use

The MSA test device Multitest plus Console HP 230 V [hereinafter referred to as test device] is designed for testing full face masks, lung governed demand valves and compressed breathing apparatus in the range of low-, medium- and high pressure.

Chemical protective suits can be tested only with special adapters.

It is imperative that this operating manual be read and observed when using the product. In particular, the safety instructions, as well as the information for the use and operation of the product, must be carefully read and observed. Furthermore, the national regulations applicable in the user's country must be taken into account for a safe use.



#### Danger!

This product is supporting life and health. Inappropriate use, maintenance or servicing may affect the function of the device and thereby seriously compromise the user's life.

Before use the product operability must be verified. The product must not be used if the function test is unsuccessful, it is damaged, a competent servicing/maintenance has not been made, genuine MSA spare parts have not been used.

Alternative use, or use outside this specification will be considered as non-compliance. This also applies especially to unauthorised alterations to the product and to commissioning work that has not been carried out by MSA or authorised persons.

### 1.2 Liability Information

MSA accepts no liability in cases where the product has been used inappropriately or not as intended. The selection and use of the product are the exclusive responsibility of the individual operator.

Product liability claims, warranties also as guarantees made by MSA with respect to the product are voided, if it is not used, serviced or maintained in accordance with the instructions in this manual.

### 1.3 Safety and Precautionary Measures

The test device is built and tested in accordance with DIN 57411 Part 1/VDE 0411 Part 1, protection measures for electronic measuring equipment and was released from the factory in a perfectly safe condition. In order to maintain this condition, and to ensure safe operation, the user must observe the instructions and warning notes which are contained in these instructions for use.

#### Connection to the Supply Voltage

Prior to switching on, please ensure that the set operating voltage and mains voltage on the test device concur. The mains connector can only be connected to a socket with sealed contact. The protective effect must not be removed by an extension without protective wire.

#### Protective Wire

Any disconnection of the protective wire, inside or outside the test device, or loosening of the protective wire connection, can make the test device dangerous. Intentional disconnection is not permitted.

#### Opening Covers

When opening covers or removing parts, live parts may be exposed. Even connection points could be live. The device must be disconnected from all voltage sources before adjustment, maintenance, corrective maintenance or the replacement of parts.

If, after adjustment, maintenance or repairs on an opened live test device are unavoidable, these may only be carried out by experts who are fully aware of the related hazards.

#### Fuses

Only the stipulated type of fuses with the given rated amperage can be used as a replacement. Do not use patched fuses or short-circuit the fuse holder.

#### Errors and Unusual Stresses

If it is ascertained that safe operation is no longer possible, the test device must be shut down and secured against unintentional switching on. Error recovery must be performed by the manufacturer's customer service or by qualified and authorised personnel.

## 2 Description

### 2.1 Scope of Delivery

- Test device Multitest plus Console - HP 230 V
- Test head with stand base, clamp and filling line
- Test line with standard thread connector for mask tightness test
- Closing cap for exhalation valve of full face masks
- Closing cap for medium pressure line of lung governed demand valve
- High pressure test line
- T-piece 300 bar
- Power cord 230 V
- Stopwatch
- Silicone spray
- User manual

### 2.2 Overview

The test device is designed for static tests of full face masks, lung governed demand valves and compressed air breathing apparatus.

Chemical protective suits can be tested only with special adapters.

The basic version of the device permits testing of negative pressure devices with standard thread connector. With suitable accessories tests can also be performed on positive pressure lung governed demand valves and masks, as well as on chemical protective suits of various manufacturers.

Following tests are possible:

- Tightness test and opening pressure of exhalation valve of masks
- Tightness test, opening pressure and closing pressure of lung governed demand valves
- Static medium pressure of pressure reducer
- Tightness tests and pressure test of the pressure reducer warning device
- Pressure gauge accuracy test
- Tightness tests and valve test of chemical protective suites

Power supply is realised by 230 V AC.

To perform the tests a compressed air connection is not required. The compressed air required to inflate the test head can be taken from a compressed air breathing apparatus or a compressed air line.

The test pressures required for the tightness and function tests are generated by the pump built into the case. The low pressure of the compressed air breathing apparatus is indicated on the low pressure gauge. The medium pressure of the compressed air breathing apparatus is indicated on the medium pressure gauge. The high pressure of the compressed air breathing apparatus is indicated on the high pressure gauge.

Inflating of chemical protective suits or the test head is done with the push button of the filling line which is equipped with a plug nipple that fits the safety plug-coupling.

The test head can be fixed on the work bench by means of a clamp.

## 2.3 Operating Elements




Fig. 1 Test Device Multitest plus Console - HP 230 V

- |   |   |    |  |
|---|---|----|--|
| 1 | ON / OFF button   | 8  | Push button for pump   |
| 2 | Pressure gauge, low pressure [+/- 30 mbar]              | 9  | Stop valve pressure pump   |
| 3 | Pressure gauge, medium pressure [16 bar]                | 10 | Standard thread connector for lung governed demand valve and test line |
| 4 | Pressure gauge, high pressure [400 bar]                 | 11 | High pressure connector for SCBA                                       |
| 5 | Medium pressure coupling for lung governed demand valve | 12 | Medium pressure connector for SCBA                                     |
| 6 | Stop valve suction pump                                 | 13 | Power cord connection [Rear Side]                                      |
| 7 | Pressure relief valve                                   |    |  |




Fig. 2 Test Head

- |   |                                     |
|---|-------------------------------------|
| 1 | Test head                           |
| 2 | Pressure relief valve for test head |
| 3 | Push button valve                   |
| 4 | Clamp                               |




Fig. 3 Test Line

- |   |  |
|---|--|
| 1 | Standard thread connector for lung governed demand valve and test line |
| 2 | Mask test line   |
| 3 | Standard thread connector for mask with peg                            |



For masks/lung governed demand valve that do not have a standard thread connector, respective adapters are required [→ chapter 9]




Fig. 4 High pressure test line

- 1 Connector for test device
- 2 High pressure test line
- 3 Connector for SCBA pressure reducer




Fig. 5 T-piece

- 1 Connector for "300 bar" cylinder
- 2 Connector for test line
- 3 Connector for SCBA pressure reducer

## 2.4 Test Head

In order to use the test head for optimum mask tests, it must be inflated to a normal size head shape [→ Fig. 6].

This is done via the filling line that has a push button valve [→ Fig. 7] which can be connected to a medium pressure supply source [e.g. medium pressure line of a compressed air breathing apparatus]. Inflating the head is done by pushing the push button.



### Attention!

If the test head is inflated excessively the rubber part can be damaged.




Fig. 6 Inflating the test head

- A Ideal head shape for mask tests
- B Test head inflated insufficiently
- C Test head inflated excessively




Fig. 7 Push button valve

- 1 Pressure relief valve
- 2 Push button
- 3 Medium pressure line

### Inflating the test head

Procedure:

- (1) Connect push button valve to medium pressure line.
- (2) Close pressure relief valve of push button valve.
- (3) Briefly activate push button to inflate the head. Observe that the head is not inflated too much.
- (4) In case it was inflated too much, reduce pressure with pressure relief valve.
- (5) Disconnect push button valve from medium pressure line.

## 2.5 Test Head Maintenance

In order to protect the test head from premature ageing inflate test head only up to normal head shape and protect against sun radiation.

In case it is not in use, keep the test head covered.

Once a week thoroughly spray the test head with silicone spray [→ chapter 9.3] and leave it on overnight [uncovered].

## 3 Use

### 3.1 Device Test

The required test criteria and intervals for tightness and function tests are specified in the operating manuals of each manufacturer and the respective European standards.



#### Attention!

The specifications in the operating manual for the devices to be tested apply.

### 3.2 Activation

To activate the test device connect the power supply and switch On the device [ON / OFF button] and the device is ready for use.

## 4 Testing Negative Pressure Devices



Close the stop valves “suction pump” and “pressure pump” before beginning any test procedure.


Otherwise your test result can be incorrect or you may damage the test device.

### 4.1 Lung Governed Demand Valve – Tightness Test without Medium Pressure



The tightness test includes a negative and a positive pressure tightness test.

#### Using Test Device Multitest plus Console HP 230 V



##### Negative Pressure Tightness Test

- (1) Connect lung governed demand valve to test connector.
- (2) Close connecting nipple of lung governed demand valve.
- (3) Open valve lever “suction pump”.
- (4) Push and hold the button for pump.
- (5) Set to slightly above required test pressure and then close valve lever.
- (6) Release button for pump.
- (7) Slowly adjust pressure to test pressure with pressure relief valve.
- (8) Measure according to manufacturer instructions.

##### Positive Pressure Tightness Test

- (1) Connect lung governed demand valve to test connector.
- (2) Close connecting nipple of lung governed demand valve.
- (3) Open valve lever “pressure pump”.
- (4) Push and hold the button for pump.
- (5) Set to slightly above required test pressure and then close valve lever.
- (6) Release button for pump.
- (7) Slowly adjust pressure to test pressure with pressure relief valve.
- (8) Measure according to manufacturer instructions.

#### 4.2 Lung Governed Demand Valve – Tightness Test with Medium Pressure

**Attention!**

Absolutely observe sequence!

Otherwise your test result can be incorrect or you may damage the test device.



- (1) Close valve levers "suction pump" and "pressure pump".
- (2) Disconnect lung governed demand valve from test connector.
- (3) Connect compressed air breathing apparatus to medium pressure connector.
- (4) Connect lung governed demand valve to medium pressure adapter [case-version] or to medium pressure connector at the test device [console].
- (5) Open cylinder[s].
- (6) Connect lung governed demand valve to test connector.
- (7) Slowly open valve lever "pressure pump".
- (8) Push and hold the button for pump.
- (9) Set to slightly above required test pressure and then close valve lever quickly.
- (10) Release button for pump.
- (11) Slowly adjust pressure to test pressure with pressure relief valve.
- (12) Measure according to manufacturer instructions.



#### 4.3 Lung Governed Demand Valve – Opening Pressure

**Attention!**

Absolutely observe sequence!

Otherwise your test result can be incorrect or you may damage the test device.



- (1) Disconnect lung governed demand valve from test connector.
- (2) Connect compressed air breathing apparatus to medium pressure adapter.
- (3) Connect lung governed demand valve to medium pressure adapter [case-version] or to medium pressure connector at the test device [console].
- (4) Open cylinder[s].
- (5) Connect lung governed demand valve to test connector.
- (6) Slowly open valve lever "suction pump".
- (7) Push and hold the button for pump.
- (8) Measure according to manufacturer instructions.




#### 4.4 Medium Pressure – Tightness Test

**Attention!**

Absolutely observe sequence!

Otherwise your test result can be incorrect or you may damage the test device.



- (1) Disconnect lung governed demand valve from test connector.
- (2) Connect compressed air breathing apparatus to medium pressure connector.
- (3) Connect lung governed demand valve to medium pressure coupling.
- (4) Open cylinder[s].
- (5) Wait until pressure has stabilised.
- (6) Close cylinder[s].
- (7) Briefly release pressure with lung governed demand valve.
- (8) Measure according to manufacturer instructions.




#### 4.5 High Pressure – Tightness Test

**Attention!**

Absolutely observe sequence!

Otherwise your test result can be incorrect, you may damage the test device or you may unintendedly set air free.



- (1) Disconnect lung governed demand valve from test connector.
- (2) Connect compressed air breathing apparatus using the high pressure line [200 bar] or the T-piece [300 bar] to high pressure connector.
- (3) Connect compressed air breathing apparatus to medium pressure connector.
- (4) Connect lung governed demand valve to medium pressure coupling.
- (5) Open cylinder[s].
- (6) Wait until pressure has stabilised.
- (7) Close cylinder[s].
- (8) Measure according to manufacturer instructions.



#### 4.6 High Pressure - Pressure and Gauge Accuracy Test



- (1) Disconnect lung governed demand valve from test connector.
- (2) Connect compressed air breathing apparatus using the high pressure line [200 bar] or the T-piece [300 bar] to high pressure connector.
- (3) Connect compressed air breathing apparatus to medium pressure connector.
- (4) Connect lung governed demand valve to medium pressure coupling.
- (5) Open cylinder[s].
- (6) Wait until pressure has stabilised.
- (7) Close cylinder[s].
- (8) Briefly release pressure with lung governed demand valve.
- (9) Measure according to manufacturer instructions.

#### 4.7 Mask - Tightness Test



- (1) Fit mask onto the test head.
- (2) Connect test line to test device and mask [connector with peg into mask].
- (3) Open valve lever "suction pump".
- (4) Push and hold the button for pump.
- (5) Set to slightly above required test pressure and then close valve lever.
- (6) Release button for pump.
- (7) Slowly adjust pressure to test pressure with pressure relief valve.
- (8) Measure according to manufacturer instructions.

## 5 Testing Positive Pressure Devices



Close the stop valves "suction pump" and "pressure pump" before beginning any test procedure.

Otherwise your test result can be incorrect or you may damage the test device.

### 5.1 Lung Governed Demand Valve – Tightness Test without Medium Pressure



#### Attention!

Switch off positive pressure.

When the positive pressure is not switched off, you will unintendedly set air free.




Choose correct Transition Adapter lung governed demand valve from chapter 9.1.



The tightness test includes a negative and a positive pressure tightness test.

Using Test Device Multitest plus Console HP 230 V



#### Negative Pressure Tightness Test

- (1) Switch off positive pressure.
  - ▷ The lung governed demand valve is in standby position.
- (2) Connect lung governed demand valve to test connector.
- (3) Close connecting nipple of lung governed demand valve.
- (4) Open valve lever "suction pump".
- (5) Push and hold the button for pump.
- (6) Set to slightly above required test pressure and then close valve lever.
- (7) Release button for pump.
- (8) Slowly adjust pressure to test pressure with pressure relief valve.
- (9) Measure according to manufacturer instructions.

**Positive Pressure Tightness Test**

- (1) Switch off positive pressure.
  - ▷ The lung governed demand valve is in standby position.
- (2) Connect lung governed demand valve to test connector.
- (3) Close connecting nipple of lung governed demand valve.
- (4) Open valve lever "pressure pump".
- (5) Push and hold the button for pump.
- (6) Set to slightly above required test pressure and then close valve lever.
- (7) Release button for pump.
- (8) Slowly adjust pressure to test pressure with pressure relief valve.
- (9) Measure according to manufacturer instructions.

**5.2 Lung Governed Demand Valve – Tightness Test with Medium Pressure****Attention!**

Absolutely observe sequence!

Otherwise your test result can be incorrect, you may damage the test device or you may unintendedly set air free.





- (1) Disconnect lung governed demand valve from test connector.
- (2) Switch off positive pressure.  
▷ The lung governed demand valve is in standby position.
- (3) Connect compressed air breathing apparatus to medium pressure connector.
- (4) Connect lung governed demand valve to medium pressure coupling at the test device.
- (5) Open cylinder[s].
- (6) Connect lung governed demand valve to test connector.
- (7) Open valve lever "pressure pump".
- (8) Push and hold the button for pump.
- (9) Set to slightly above required test pressure and then close valve lever quickly.
- (10) Release button for pump.
- (11) Slowly adjust pressure to test pressure with pressure relief valve.
- (12) Measure according to manufacturer instructions.

### 5.3 Lung Governed Demand Valve – Closing Pressure



- (1) Disconnect lung governed demand valve from test connector.
- (2) Switch on positive pressure.  
▷ The lung governed demand valve is in positive pressure position.
- (3) Close connector of lung governed demand valve with the palm of the hand.
- (4) Connect lung governed demand valve to medium pressure adapter [case-version] or to medium pressure connector at the test device [console].
- (5) Connect compressed air breathing apparatus to medium pressure adapter.
- (6) Open cylinder[s].  
▷ Exhaust port in direction of transition adapter.
- (7) Connect lung governed demand valve to transition adapter.
- (8) Briefly release pressure with pressure relief valve for about 15 s.
- (9) After 15 s close the pressure relief valve.
- (10) Measure according to manufacturer instructions.





## 5.4 Medium Pressure – Tightness Test

**Attention!**

Absolutely observe sequence!

Otherwise your test result can be incorrect or you may damage the test device.



- (1) Disconnect lung governed demand valve from test connector.
- (2) Switch off positive pressure.
  - ▷ The lung governed demand valve is in standby position.
- (3) Connect lung governed demand valve to medium pressure adapter [case-version] or to medium pressure coupling at the test device [console].
- (4) Connect compressed air breathing apparatus to medium pressure adapter [case-version] or to medium pressure coupling at the test device [console].
- (5) Connect medium pressure adapter to test connector.
- (6) Open cylinder[s].
- (7) Wait until pressure has stabilised.
- (8) Close cylinder[s].
- (9) Briefly release pressure with lung governed demand valve.
- (10) Measure according to manufacturer instructions.


## 5.5 High Pressure – Tightness Test



### Attention!

Absolutely observe sequence!

Otherwise your test result can be incorrect, you may damage the test device or you may unintendedly set air free.



- (1) Disconnect lung governed demand valve from test connector.
- (2) Switch off positive pressure.
  - ▷ The lung governed demand valve is in standby position.
- (3) Connect compressed air breathing apparatus using the high pressure line [200 bar] or the T-piece [300 bar] to high pressure connector.
- (4) Connect compressed air breathing apparatus to medium pressure connector.
- (5) Connect lung governed demand valve to medium pressure coupling.
- (6) Open cylinder[s].
- (7) Wait until pressure has stabilised.
- (8) Close cylinder[s].
- (9) Measure according to manufacturer instructions.

## 5.6 High Pressure - Pressure and Gauge Accuracy Test



- (1) Disconnect lung governed demand valve from test connector.
- (2) Switch off positive pressure.
  - ▷ The lung governed demand valve is in standby position.
- (3) Connect compressed air breathing apparatus using the high pressure line [200 bar] or the T-piece [300 bar] to high pressure connector.
- (4) Connect compressed air breathing apparatus to medium pressure connector.
- (5) Connect lung governed demand valve to medium pressure coupling.
- (6) Open cylinder[s].
- (7) Wait until pressure has stabilised.
- (8) Briefly release pressure with lung governed demand valve.
- (9) Measure according to manufacturer instructions.

## 5.7 Mask - Tightness Test



Choose correct Transition Adapter lung governed demand valve from chapter 9.1.



- (1) Fit mask onto the test head.
- (2) Connect test line to test device and mask [connector with peg into mask].
- (3) Open valve lever “suction pump”.
- (4) Push and hold the button for pump.
- (5) Set to slightly above required test pressure and then close valve lever.
- (6) Release button for pump.
- (7) Slowly adjust pressure to test pressure with pressure relief valve.
- (8) Measure according to manufacturer instructions.

## 5.8 Mask - Opening Pressure of Exhalation Valve



- (1) Fit mask onto the test head.
- (2) Connect test line to test device and mask [connector with peg into mask].
- (3) Open valve lever “pressure pump”.
- (4) Push and hold the button for pump.
- (5) Measure according to manufacturer instructions.

## 6 Testing Chemical Protective Suit



For testing chemical protective suits special accessories may be necessary.

For detailed information see chapter 9.2 and operating manual of the manufacturer.

### 6.1 Preparation



- (1) Spread out chemical protective suit on suitable place.
  - e.g. work bench or clean floor



- (2) Remove all angled prechambers.
- (3) Remove all exhalation valve discs.
- (4) Connect test line to exhalation valve.
- (5) Connect test line with test device.
- (6) Connect filling line.



- (7) Close all remaining exhalation valves.

### 6.2 Stabilising the Chemical Protective Suit



#### Attention!

Do not exceed the required stabilizing pressure since the seams and the glued sections of the suit may be damaged.



- (1) Connect test line to test device.
- (2) Connect push button valve of filling line to a clean compressed air source
  - e.g. compressed air breathing apparatus, pressure reducer, etc.
- (3) Set the stabilising pressure with the push button valve.
  - Stabilise according to manufacturer instructions.
- (4) Measure according to manufacturer instructions.

### 6.3 Chemical Protective Suit – Tightness Test

**Attention!**

Stabilisation of the chemical protective suit must have been performed prior to the tightness test!

Otherwise the test result is incorrect.



- (1) Disconnect push button valve of the filling line of the air supply source.
- (2) Reduce pressure in the chemical protective suit according to manufacturer instructions to the required test pressure with the push button valve of the filling line.
- (3) Measure according to manufacturer instructions.

### 6.4 Chemical Protective Suit – Valve Test

**Attention!**

The following procedure must be performed for every single suit valve!



The valves must be ready for use.



- (1) Close valve of the chemical protective suit from the inside with plug for the valve tightness test.
- (2) Connect test line to exhalation valve.
- (3) Connect test line to test device.
- (4) Open valve lever "suction pump".
- (5) Push and hold the button for pump.
- (6) Set to slightly above required test pressure and then close valve lever.
- (7) Release button for pump.
- (8) Slowly adjust pressure to test pressure with pressure relief valve.
- (9) Measure according to manufacturer instructions.

## 7 Test Criteria for MSA Respiratory Protection Apparatus



The test criteria given below serve only as basis for evaluation.

The actual required test criteria are specified in the instruction manuals of the manufacturers!

Apparatus	Test	Test Criteria	Remarks
<b>Negative pressure lung governed demand valve</b>	Tightness test with positive and negative pressure	At $\pm 7.5$ mbar, pressure change 1.0 mbar in 1 minute	Lung governed demand valve unpressurised
	Opening pressure [initiating pressure]	< -3.5 mbar	Lung governed demand valve with medium pressure 10 bar
<b>Positive pressure lung governed demand valve</b>	Tightness test with positive pressure	At +7.5 mbar, pressure change 1.0 mbar in 1 minute	Lung governed demand valve unpressurised
	Closing pressure	2.0 mbar to 3.9 mbar	Lung governed demand valve with medium pressure in positive pressure position
<b>Negative pressure full face mask</b>	Tightness test mask with exhalation valve with negative pressure	At -10 mbar maximum 1 mbar pressure change in 1 minute	Valve disc moist
<b>Positive pressure full face mask</b>	Tightness test with mask and exhalation valve with negative pressure	At -10 mbar maximum 1 mbar pressure change in 1 minute	Valve disc moist
	Opening pressure exhalation valve	> 4.2 mbar	
<b>Chemical protective suit</b>	Tightness test with positive pressure	At 16 mbar maximum 2 mbar pressure change in 3 minutes	Filling pressure for stabilizing 18 mbar. For Champion 25 mbar
	Tightness test suit valves with negative pressure	At -10 mbar maximum 1 mbar pressure change in 1 minute	Valve disc moist

## 8 Technical Data

<b>Multitest plus Console - HP 230 V</b>	
Dimensions [W x H x D] in mm	420 x 210 x 400
Power Supply	230 V/50 Hz, 100 W
Weight	9 kg
Pressure gauge measuring range	
- Low pressure	± 30 mbar
- Medium pressure	0 – 16 bar
- High pressure	0 – 400 bar
Pressure gauge precision class	
- Low pressure	1.6 %
- Medium pressure	1.0 %
- High pressure	0.6 %



## 9 Ordering Information

### 9.1 Transition Adapters Mask / Lung Governed Demand Valves


Description	Part No.		
		Lung governed demand valve [1]	Mask [2]
Transition Adapter MSA AUER LA96-AS	D4080891		D5175536
Transition Adapter MSA AUER AutoMaXX	10031901		10089532
Transition Adapter Dräger Steck	D5175512		D5135537
Transition Adapter M45 x 3	D4074833		D5175538
Transition Adapter ESA	10037962		

**9.2 Transition Adapter for Chemical Protective Suit**

Description	Part No.
Transition Adapter for Chemical Protective Suit – Dräger	D5175521
Transition Adapter for Chemical Protective Suit – Koch	D5175522
Transition Adapter for Chemical Protective Suit - Trelleborg	D5175523

## 9.3 Accessories

Description	Part No.
<b>Test Line</b> for testing chemical protective suits	D5175532
	
<b>Filling Line</b> for testing chemical protective suits	D5175518
	
<b>MSA AUER sealing cap</b> to close valves on Auer chemical protective suits and 3S negative pressure masks	D5135047
	
<b>Valve test device</b> for MSA AUER chemical protective suits	D5175533
	
<b>Barcode Scanner</b>	10025433
	

Description	Part No.
<b>Barcode Labels</b> for using inside [masks] or outside [SCBA or cylinders] 100 pieces	
	
Barcode Labels – outside	10025420
Barcode Labels – inside	10025422
<b>Test Gauge</b> For testing pressure of compressed air cylinders 200 bar and 300 bar.	D4080929
	
<b>Silicone spray</b>	10017461
<b>Disinfectant</b> For disinfecting instruments, masks and others.	on request
<b>Ultrasonic Washing Machine</b> for 9 masks	D5175709
	

# MSA in Europe

[ [www.msa-europe.com](http://www.msa-europe.com) & [www.msa-gasdetection.com](http://www.msa-gasdetection.com) ]

## Northern Europe

### Netherlands

**MSA Nederland**  
Kernweg 20  
1627 LH Hoorn  
Phone +31 [229] 25 03 03  
Fax +31 [229] 21 13 40  
info@msaned.nl

### Belgium

**MSA Belgium**  
Duwijkstraat 17  
2500 Lier  
Phone +32 [3] 491 91 50  
Fax +32 [3] 491 91 51  
msabelgium@msa.be

### Great Britain

**MSA Britain**  
East Shawhead  
Coatbridge ML5 4TD  
Scotland  
Phone +44 [12 36] 42 49 66  
Fax +44 [12 36] 44 08 81  
info@msabritain.co.uk

### Sweden

**MSA NORDIC**  
Kopparbergsgatan 29  
214 44 Malmö  
Phone +46 [40] 699 07 70  
Fax +46 [40] 699 07 77  
info@msanordic.se

### MSA SORDIN

Rörläggervägen 8  
33153 Värnamo  
Phone +46 [370] 69 35 50  
Fax +46 [370] 69 35 55  
info@sordin.se

## Southern Europe

### Italy

**MSA Italiana**  
Via Po 13/17  
20089 Rozzano [MI]  
Phone +39 [02] 89 217 1  
Fax +39 [02] 82 59 228  
info-italy@  
msa-europe.com

### Spain

**MSA Española**  
Narcís Monturiol, 7  
Pol. Ind. del Sudoeste  
08960 Sant-Just Desvern  
[Barcelona]  
Phone +34 [93] 372 51 62  
Fax +34 [93] 372 66 57  
info@msa.es

### France

**MSA GALLET**  
Zone Industrielle Sud  
01400 Châtillon sur  
Chalaronne  
Phone +33 [474] 55 01 55  
Fax +33 [474] 55 47 99  
message@msa-gallet.fr

## Eastern Europe

### Poland

**MSA Safety Poland**  
ul. Wschodnia 5A  
05-090 Raszyn k/Warszawy  
Phone +48 [22] 711 50 33  
Fax +48 [22] 711 50 19  
mee@msa-europe.com

### Czech Republic

**MSA Safety Czech**  
Pikartská 1337/7  
716 07 Ostrava-Radvanice  
Phone +420 [59] 6 232222  
Fax +420 [59] 6 232675  
info@msa-auer.cz

### Hungary

**MSA Safety Hungaria**  
Francia út 10  
1143 Budapest  
Phone +36 [1] 251 34 88  
Fax +36 [1] 251 46 51  
info@msa-auer.hu

### Romania

**MSA Safety Romania**  
Str. Virgil Madgearu, Nr. 5  
Ap. 2, Sector 1  
014135 Bucuresti  
Phone +40 [21] 232 62 45  
Fax +40 [21] 232 87 23  
office@msanet.ro

### Russia

**MSA Russia**  
Leninsky Prospect 2  
9th Floor, office 14  
119049 Moscow  
Phone +7 [495] 544 93 89  
Fax +7 [495] 544 93 90  
msa-russia@  
msa-europe.com

## Central Europe

### Germany

**MSA AUER**  
Thiemannstrasse 1  
12059 Berlin  
Phone +49 [30] 68 86 0  
Fax +49 [30] 68 86 15 17  
info@auer.de

### Austria

**MSA AUER Austria**  
Kaplanstrasse 8  
3430 Tulln  
Phone +43 [22 72] 63 360  
Fax +43 [22 72] 63 360 20  
info@msa-auer.at

### Switzerland

**MSA Schweiz**  
Eichweg 6  
8154 Oberglatt  
Phone +41 [43] 255 89 00  
Fax +41 [43] 255 99 90  
info@msa.ch

### European International Sales

[Africa, Asia, Australia, Latin America, Middle East]

### MSA EUROPE

Thiemannstrasse 1  
12059 Berlin  
Phone +49 [30] 68 86 55 5  
Fax +49 [30] 68 86 15 17  
contact@msa-europe.com