

# MET LABORATORIES, INC. CERTIFICATION RECORD



*The applicant named below has been authorized by MET Laboratories, Inc. to represent the product(s) listed in this record as "MET Certified" and to mark this/these product(s) according to the terms and conditions of the MET Mark Utilization Agreement, MET Listing Reports, and the applicable marking agreements. Only the product(s) bearing the MET Mark and under a follow-up service are considered to be included in the MET Certification program.*

**FILE NUMBER:** E112409

**APPROVAL DATE:** May 11, 2005

**REVISED:** -

| PRODUCT(S)                    | MODEL(S)           | ELECTRICAL RATINGS  |
|-------------------------------|--------------------|---|
| Gas Monitor (Metal Enclosure) | Ultima X Series    | 12Vdc Supply: 100-240 VAC, 50-60Hz, 15W<br>24Vdc Supply: 00-240 VAC, 50-60 Hz, 7W |
| Sensors                       | Ultima XE and X-IR | Powered from the gas monitor unit.  |

| STANDARD NUMBER                | STANDARD TITLE  | EDITION |
|--------------------------------|---|---------|
| UL61010-1/CSA C22.2 No.61010-1 | Electrical Equipment For Measurement, Control, and Laboratory Use | Second  |

MET LABORATORIES, INC. requires that any and all changes proposed in the previously identified product(s), that affects the information contained in the above referenced listing report, must be submitted to MET for evaluation prior to implementation to assure continued MET Certification status.

The above identified product(s) has/have been submitted by the applicant:

**APPLICANT:**

Mine Safety Appliances  
Instrument Division  
P.O. Box 427  
Pittsburgh, PA 15230 USA

The covered products shall be subjected to semi annual follow-up inspections to ensure that the Certified product(s) are identical to the representative product sample evaluated by MET LABORATORIES, INC. and that all manufacturer's responsibilities are being fulfilled as specified in the MANUFACTURING RESPONSIBILITY section of the Certification report.

Michael Terebuch  
Project Engineer  
Safety Laboratory

Rick Cooper  
Director of Laboratory Operations,  
Safety Laboratory



*MET Laboratories, Inc. is accredited by OSHA and the Standards Council of Canada.  
The Nation's First Nationally Recognized Testing Laboratory*

