D-Fend™ Water Traps

Change-out guide



Why should you change water traps following the instructions below?

D-Fend range water trap protects your GE Monitor against **humidity and contamination**

It also acts as a dust and **bacteria/virus filter** to safeguard measurement sensors inside the respiratory module

Contamination may **affect module performance** and may lead to module failures

Keep D-Fend Pro water trap inserted in the module at all time including during **transport and storage** to protect sensors inside

> *To ease the water trap's change-out, mark the date on the cartridge when placed in use.*

D-Fend water traps help protect your GE Healthcare respiratory monitoring equipment from **humidity, secretions, viral and bacterial contamination**. The D-Fend water trap is indicated for use with respiratory modules in both the anesthesia and critical care environments.

Bacteriel and Viral Filtration Efficiency

PART	BFE	VFE
D-Fend D-Fend+ Mini D-Fend	>99,99998%	99,99997 - 99,99999%
D-Fend Pro D-Fend Pro+	>99,999970%	>99,9988%



There are application-specific versions for both Anesthesia and Critical Care environments. Note that the expected service life of water traps for critical care is **24 hours** due to more humid environmental conditions caused by active humidification, patient cooling or heating devices.

Mini D-Fend™ Mini D-Fend water trap is specified for use with the following module: **E-miniC**



Mini D-Fend - Grey — **8002174** — 10 / Case

In anesthesia environment, the maximum lifetime of a water trap is two months.
In Critical Care environment, change water trap for each new patient, or every 24 hours

D-Fend™ Pro and D-Fend Pro+ water traps

D-Fend Pro and D-Fend Pro+ water traps for single-width CARESCAPE™ respiratory modules



D-Fend Pro - *for use in **ANESTHESIA*** - Dark Steel Blue — **M1182629** — 10 / Case

The maximum lifetime of a water trap is two months



D-Fend Pro+ - *for for use in **ICU*** - Green — **M1200227** — 10 / Case

Change water trap for each new patient, or at least every 24 hours

D-Fend™ water traps D-Fend water traps for double-width compact airway gas modules



D-Fend - *for use in **ANESTHESIA*** - Black — **876446-HEL** — 10 / Case

The maximum lifetime of a water trap is two months



D-Fend+ - *for for use in **ICU*** - Green — **88139-HE** — 10 / Case

Change water trap for each new patient, or at least every 24 hours

Scan me and view
How-to change water trap video

gehealthcare.co.uk

Data subject to change. GE and GE Monogram are trademarks of General Electric Company.™ Trademarks of General Electric Company. ©2021 General Electric Company. All rights reserved. JB04534XE