

3M[™] Tegaderm[™] PICC/CVC Securement System 1837-2100 CVC – Internal jugular

Removal

Application



Apply 3M[™] Cavilon[™] No Sting Barrier Film around the insertion site where the dressing will adhere. Avoid the area within 2cm of the catheter insertion site. Allow all preps and skin protectants to dry completely. Open package and remove sterile dressing and device. Put the 'Post-It' application guide somewhere visible.



Orient the device with arrows pointing toward the insertion site. Place single-, double- or triple lumen catheters into the device securement post and weave lumen(s) under the single plastic arm.



Remove the liner of the attached tape strip and secure lumen(s) to the device base.



Position device on skin at the desired location. Pull and remove liner from one side of base to expose adhesive while holding the device in place. Pull and remove liner from the other side of base. Reposition the device if necessary to gain correct position.



Remove documentation label. Lift catheter lumens, exposing the notch of the perforated tape strip. Gently pull apart the perforated tape strip either side of the catheter.



Utilising the low and slow removal technique, slowly start removing the tape strips and dressing, as one layer, toward the insertion site until the dressing is beyond the securement device.



Peel the backing from the dressing, exposing the adhesive surface. Orient the adhesive to the skin and place the dressing so the transparent film covers the insertion site and the flat border edge of the dressing covers the plastic base of the device. Do not stretch dressing during application.



Remove the frame delivery system whilst applying firm pressure to the dressing and dressing edges to enhance adhesion to the skin. Place the frame to the side.



Remove the backing from securement strip. Lift the catheter lumen(s) and apply the notched end of the securement strip under the catheter lumen(s) and over the dressing edge ensuring all of the securement device is covered. Apply pressure to enhance adhesion. Remove the frame delivery system.



Document dressing change on the documentation label. Place the documentation label on top of the dressing over the catheter lumen(s) to close the system. Apply pressure to enhance adhesion.



Remove the device tape strip from the catheter lumen(s) and stick to one side to prevent intrusion. Unhook the catheter from under the fixed arm and from out of the securement post.



Secure the catheter with one hand and use your other hand to remove the device from the patient's skin. If the device is to be replaced, clean the skin, allow to dry and apply a new device. Once secure, remove the remaining dressing 'low & slow', clean the area and apply a new dressing.

3M United Kingdom PLC Charnwood Campus 10 Bakewell Road Loughborough LE11 5RB 01509 611611 3M Ireland The Iveagh Building Carrickmines Park Carrickmines Dublin 18, Ireland 00 353 (01) 280 3555

www.3M.co.uk/vascularaccess

GF511. J360555. 3M and Tegaderm are trademarks of the 3M company. © 3M 2017. All rights reserved.