

3M™ Sterilization Assurance Quick Reference Guide



For 3M™ Attest™ Auto-reader 390* and 3M™ Attest™ Rapid Readout Biological Indicators (BIs)**

1	Cool		Allow test BI to cool at least 10 minutes.
2	Close		While wearing safety glasses, close the cap of the cooled BI by pushing the cap down.
3	Crush		Hold BI by cap and use “crusher well” to crush the media ampoule.
4	Tap		Holding BI by cap, tap the bottom of the vial on the counter until media wets spore strip.
5	Place		Place the activated BI in color-coded incubation well.
6	Interpret Result		Observe LCD panel under incubating BIs for negative (-) or positive (+) result. The positive control should read positive (+). A positive result for a test (processed) BI, which indicates a sterilization process failure, should be acted upon immediately.
7	Record		Record result for both test BI (processed) and positive control BI (unprocessed).

Optional Visual pH Color Change Result



Negative

- Media remains purple.
- A negative pH color change result can be assessed at 24 hours for Attest™ biological indicators 1291 and at 48 hours for Attest™ biological indicators 1292.



Positive

- A media color change from purple to yellow indicates a positive result.
- The biological indicator may be discarded as soon as the positive result has been observed and documented.



Infection Prevention Division
3M Health Care
2510 Conway Avenue
St. Paul, MN 55144-1000
U.S.A.
1 800 228-3957
www.3M.com/infectionprevention

Please recycle. Printed in U.S.A.
3M and Attest are trademarks of 3M.
© 3M 2013. All rights reserved.
70-2010-9397-1

* Refer to 3M™ Attest™ Auto-reader 390 Operator's Manual for complete instructions
** Refer to 3M™ Attest™ Rapid Readout Biological Indicator (BI) Instructions for Use for complete instructions