



Curoos™

Disinfecting Port
Protectors

How to use

3M™ Curoos™


Disinfecting Cap for Needleless Connectors

Disinfects in one minute – protects needleless connectors for up to seven days.



Peel

the foil from the 3M™ Curoos™ Disinfecting Cap.



Push and twist

the Curoos disinfecting cap clockwise onto the end of the needleless connector.

- ▶ If needleless connector is heavily soiled, clean with 70% isopropyl alcohol swab before using cap
- ▶ Do not use cap if seal is broken, torn, punctured, or sponge appears dry
- ▶ Cap must remain on needleless connector for a minimum of one minute and may remain on for seven days if not removed
- ▶ The needleless connector can be accessed immediately after removing the Curoos disinfecting cap that has been in place for a minimum of one minute
- ▶ Always place a new Curoos disinfecting cap on needleless connector after each use
- ▶ Dispose of the Curoos disinfecting cap after every use
- ▶ Follow hospital protocol for disinfecting needleless connectors, if performing sequential accesses

Warning: to avoid potential for injury, use only on needleless connectors.

Caution: potential choking hazard.