

Product		Cat. no.	Packaging	Indications for use	Time to result	Incubator/auto-reader
Super Rapid Readout		1491 Dark blue cap	50 each/box 4 boxes/case	 Gravity: 270°F (132°C) = 3 or 10 minutes 275°F (135°C) = 3 or 10 minutes		 490/490M Auto-readers
		1492V Brown cap	50 each/box 4 boxes/case	 Dynamic-air-removal (pre-vacuum and SFPP): 270°F (132°C) = 3, 3.5, 4, 5.5 or 6 minutes 273°F (134°C) = 3 or 4 minutes 275°F (135°C) = 3, 3.5 or 10 minutes		
		1493 Light blue cap	50 each/box 4 boxes/case	 Gravity: 250°F (121°C) = 30 minutes Dynamic-air-removal (pre-vacuum and SFPP): 250°F (121°C) = 15, 20, 30 or 35 minutes		
		1493P Light blue cap	20 each/box 5 boxes/case			
		41482V	24 packs/box 24 controls/box	 Dynamic-air-removal (pre-vacuum and SFPP): 270°F (132°C) = 4 minutes 275°F (135°C) = 3 minutes		
		41482VF	24 packs/box 5 controls/box			
		1492PCD-24	24 packs 24 controls/case	 Dynamic-air-removal (pre-vacuum and SFPP): 270°F (132°C) = 4 minutes 273°F (134°C) = 4 minutes 275°F (135°C) = 3 minutes		
		1492PCD-05	24 packs 5 controls/case			
	1493PCDG-24	24 packs 24 controls/case	 Gravity displacement: 250°F (121°C) = 30 minutes			
	1492PCDE-24	24 packs 24 controls/case	 Dynamic-air-removal: 270°F (132°C) = 10 minutes			
Rapid Readout		1291	50 each/box 4 boxes/case	 Gravity: 270°F (132°C) ≥ 3 minutes		 390 Auto-reader
		1292	50 each/box 4 boxes/case	 Gravity: 250°F (121°C) ≥ 20 minutes Vacuum-assisted: 270°F (132°C) ≥ 4 minutes		
		1296	25 packs/box 25 controls/box	 Gravity: 250°F (121°C) ≥ 40 minutes Vacuum-assisted: 270°F (132°C) ≥ 4 minutes		
		41382	16 packs/box 16 controls/box	 Gravity: 250°F (121°C) = 30 minutes Dynamic-air-removal 270°F (132°C) = 4 minutes		
		41382F	16 packs/box 4 controls/box			
Standard		1261	100 each/box 4 boxes/case	 Gravity: 270°F (132°C)		 116 Incubator
		1262	100 each/box 4 boxes/case	 Gravity: 250°F (121°C) Vacuum-assisted: 270°F (132°C) Gravity (table-top sterilizers only): 270°F (132°C) > 3 minutes		
		1262P	25 each/box 4 boxes/case			

*3M™ Attest™ Biological Indicators 1491, 1492V, 1493 can be read in 24 minutes when used with 3M™ Attest™ Mini Auto-reader 490M or with 3M™ Attest™ Auto-reader 490 or 490H with software version 4.0.0 or greater.

**Must be used in conjunction with 3M™ Attest™ Biological Indicator Incubator 116.